

CERTIFICATE OF ANALYSIS

Page 1 of 1

CUSTOMER

LiViliti Health Products
2140 SW Main Blvd
Lake City FL 32025

CERTIFICATE NO.

1042648.125/15749

CUSTOMER REF.

210527

SAMPLE DETAILS

LIVILITI HEALTH PRODUCTS

DATE RECEIVED

09/07/2021

ORDER NO.

METHOD: Determination of Antibacterial Activity using Test Based on MOD ISO 22196

DATE ANALYSED

14/07/2021

DATE REPORTED

16/07/2021

RESULTS (AS CFU CM⁻²)

SAMPLE	SPECIES	CONTACT TIME		REDUCTION (INITIAL)	
		0 hrs	24 hrs	Log 10	%
LIVILITI HEALTH PRODUCTS ANTIMICROBIAL TUBING	<i>E. coli</i>	1.1E+04	3.4E+05		
	<i>E. coli</i>	1.1E+04	< 1.00	≥ 4.02	≥ 99.99%
LIVILITI HEALTH PRODUCTS ANTIMICROBIAL TUBING	<i>S. aureus</i>	2.1E+04	1.4E+04		
	<i>S. aureus</i>	2.1E+04	< 1.00	≥ 4.32	≥ 99.99%

Key: NS = Poor survival on control supplied.

The above data show the difference in the population following contact with the surface of the samples listed for 24 hours at 35°C under a RH of > 95% relative to the initial population.

Certificate valid for up to 2 years from date of test.

IMSL MICROBIOLOGICAL SERVICES LTD
PALE LANE
HARTLEY WINTNEY
HANTS RG27 8DH
UK

MANAGING DIRECTOR
Peter D Askew

